



NEILSON RESEARCH CORPORATION

Environmental Testing Laboratory

4/11/2017

Dale Giovannetti
Central Point Sch. Dist/Sams Valley Elem
300 Ash St.
Central Point, OR 97502

TEL: (541) 494-6924
FAX

RE: Lead Testing - Patrick

Order No.: 1704077

Dear Dale Giovannetti:

Neilson Research Corporation received 1 sample(s) on 4/4/2017 for the analyses presented in the following report.

The results relate only to the parameters tested or to the sample as received by the laboratory. This report shall not be reproduced except in full, without the written approval of Neilson Research Corporation. If you have any questions regarding these test results, please feel free to call.

Sincerely,
Neilson Research Corporation

Ronald W. Canady
Project Manager

245 S Grape St • Medford, OR 97501 • (541) 770-5678
400 SE G St, Suite B • Grants Pass, OR 97526 • (541) 479-4053

www.nrclabs.com

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

CLIENT: Central Point Sch. Dist/Sams Valley Elem
Project: Lead Testing - Patrick
Lab Order: 1704077

Date: 11-Apr-17

CASE NARRATIVE

The analyses were performed according to the guidelines in the Neilson Research Corporation Quality Assurance Program. This report contains analytical results for the sample(s) as received by the laboratory.

Neilson Research Corporation certifies that this report is in compliance with the requirements of NELAP. No unusual difficulties were experienced during analysis of this batch except as noted below or qualified with data flags on the reports.

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Central Point Sch. Dist/Sams Valley Ele

300 Ash St.

Central Point, OR 97502

Sample Information: Lead Testing - Patrick

Lab Order: 1704077

Received Date: 4/4/2017 8:48:00 AM

Reported Date: 4/11/2017 10:00:45 AM

Lab ID: 1704077-01

Collection Date: 3/31/2017 8:45:00 AM

Matrix: DRINKING WATER

Client Sample ID: Bottle #26838

Source: City Water

Sample Location: Rm 11 Drinking Fountain

Trace Metals by EPA 200.8 ICP-MS						Analyst: OML	NELAP
Analyses	Result	Qual	MRL	Units	Dilution Factor	Date Analyzed	Accredited
Lead	0.00583		0.0001	mg/L	1	4/7/2017	A

Qualifiers:

* Value exceeds Maximum Contaminant Level
E Value above quantitation range
J Analyte detected below quantitation limits
S Spike Recovery outside accepted recovery limits

B Analyte detected in the associated Method Blank
H Holding times for preparation or analysis exceeded
ND Not Detected at the Minimum Reporting Limit

CLIENT: Central Point Sch. Dist/Sams Valley Elem
Work Order: 1704077
Project: Lead Testing - Patrick

ANALYTICAL QC SUMMARY REPORT

TestCode: ICPMS_200.8_SCHOOL

Sample ID MB-37985	SampType: MBLK	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 4/5/2017	RunNo: 94341						
Client ID: ZZZZZ	Batch ID: 37985	TestNo: EPA 200.8	(EPA 200.8)	Analysis Date: 4/7/2017	SeqNo: 1429618						
Analyte	Result	MRL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Lead ND 0.000100

Sample ID LCS-37985	SampType: LCS	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 4/5/2017	RunNo: 94341						
Client ID: ZZZZZ	Batch ID: 37985	TestNo: EPA 200.8	(EPA 200.8)	Analysis Date: 4/7/2017	SeqNo: 1429619						
Analyte	Result	MRL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Lead 0.09931 0.000100 0.1 0.000012 99.3 85 115

Sample ID 1704130-01AMS	SampType: MS	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 4/5/2017	RunNo: 94341						
Client ID: ZZZZZ	Batch ID: 37985	TestNo: EPA 200.8	(EPA 200.8)	Analysis Date: 4/7/2017	SeqNo: 1429633						
Analyte	Result	MRL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Lead 0.09299 0.000100 0.1 0.000019 93.0 70 130

Sample ID 1704130-01AMSD	SampType: MSD	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 4/5/2017	RunNo: 94341						
Client ID: ZZZZZ	Batch ID: 37985	TestNo: EPA 200.8	(EPA 200.8)	Analysis Date: 4/7/2017	SeqNo: 1429634						
Analyte	Result	MRL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Lead 0.09317 0.000100 0.1 0.000019 93.2 70 130 0.09299 0.196 20

Qualifiers: E Value above quantitation range H Holding times for preparation or analysis exceeded J Analyte detected below quantitation limits
 ND Not Detected at the Minimum Reporting Limit R RPD outside accepted recovery limits S Spike Recovery outside accepted recovery limits



Environmental Testing Laboratory
 245 South Grape Street • Medford, OR 97501
 (541) 770-6678 fax (541) 770-2901

Chain of Custody Record

This Chain of Custody is a LEGAL DOCUMENT and must be filled out accurately.

Section A

Required Client Information

Company: Central Point School District #1
 Address: 300 Ash Street
 Central Point OR 97502
 Email:
 Phone:
 Fax:
 Collected By (Print):
 Collected By (Sign):
 Email Report Yes No Mail Report Yes No
 Fax Report Yes No

Section B

Required Project Information

Project Name: Patricia
 Project Number:
 Report To:
 Copy To:
 Attention:
 Company Name:
 Address:
 P.O. #

Section C

Invoice Information

Analysis Requested

Section D

Rush Status (Subject to Scheduling)

Standard 10-14 Days
 5 Business Days (50% surcharge)
 3 Business Days (75% surcharge)
 24 - 48 hours (100% surcharge)
 Other _____
 Authorized Yes No

Section E

Sample Information

Sample ID	Comp/Grab	Matrix*	Date Collected	Time Collected	No. of Containers
RM 11 Drinking Fountain	26938	DW	3-31-17	8:45 AM	1

Section F
 Requisition/Receive
 Matrix: DW - Drinking Water WW - Wastewater W - Water S - Soil/Solid SL - Sludge O - Oil WP - Wipe OT - Other
 Requisitioned By: Sign
 Received By: Self car
 Relinquished By:
 Relinquished By:
 Received By Laboratory: Denise Reed
 Matrix: DW - Drinking Water WW - Wastewater W - Water S - Soil/Solid SL - Sludge O - Oil WP - Wipe OT - Other

Section G

Lab Use Only

NRC Workorder # 1701077
 (Lab Use Only)
 Remarks/Field Data
 NRC Sample # (Lab Use Only) 01
 Temp: AWD
 4°C +/- 2°C: Yes ___ No ___
 Received on Ice: Yes ___ No ___
 Number of Bottles Received: 1
 pH Checked: NA
 COC Seals Intact: Yes ___ No ___ NA ___
 Field Blank included: Yes ___ No ___
 Received Via: UPS FedEX Other Hand
 Payment: Invoice Cash VISA MC Check # Amount