



NEILSON RESEARCH CORPORATION

Environmental Testing Laboratory

2/23/2018

Dale Giovannetti
Central Point Sch. Dist/Sams Valley Elem
300 Ash St.
Central Point, OR 97502

TEL: (541) 494-6924
FAX

RE: Lead Testing-D.O.

Order No.: 1802626

Dear Dale Giovannetti:

Neilson Research Corporation received 3 sample(s) on 2/16/2018 for the analyses presented in the following report.

The results relate only to the parameters tested or to the sample as received by the laboratory. This report shall not be reproduced except in full, without the written approval of Neilson Research Corporation. If you have any questions regarding these test results, please feel free to call.

Sincerely,
Neilson Research Corporation

Tamra R. Schmedemann
Project Manager

245 S Grape St • Medford, OR 97501 • (541) 770-5678
400 SE G St, Suite B • Grants Pass, OR 97526 • (541) 479-4053

www.nrclabs.com

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR0028

CLIENT: Central Point Sch. Dist/Sams Valley Elem
Project: Lead Testing-D.O.
Lab Order: 1802626

Date: 23-Feb-18

CASE NARRATIVE

The analyses were performed according to the guidelines in the Neilson Research Corporation Quality Assurance Program. This report contains analytical results for the sample(s) as received by the laboratory.

Neilson Research Corporation certifies that this report is in compliance with the requirements of NELAP. No unusual difficulties were experienced during analysis of this batch except as noted below or qualified with data flags on the reports.

The EPA recommended action level for lead in schools is 0.020 mg/L.

Neilson Research Corporation

245 South Grape Street, Medford, Oregon 97501 541-770-5678 Fax 541-770-2901

Analysis Report

ORELAP 100016
EPA OR00028

Central Point Sch. Dist/Sams Valley Ele

300 Ash St.

Central Point, OR 97502

Sample Information: Lead Testing-D.O.

Lab Order: 1802626

Received Date: 2/16/2018 9:17:00 AM

Reported Date: 2/23/2018 12:16:15 PM

Lab ID: 1802626-01

Collection Date: 2/16/2018 6:26:00 AM

Matrix: DRINKING WATER

Client Sample ID: Staffroom
Source: City Water
Sample Location: Staffroom

Trace Metals by EPA 200.8 ICP-MS					Dilution	Analyst: CSB	NELAP
Analyses	Result	Qual	MRL	Units	Factor	Date Analyzed	Accredited
Lead	0.00605		0.0001	mg/L	1	2/20/2018	A

Lab ID: 1802626-02

Collection Date: 2/16/2018 6:22:00 AM

Matrix: DRINKING WATER

Client Sample ID: Tech-Sink
Source: City Water
Sample Location: Tech-Sink

Trace Metals by EPA 200.8 ICP-MS					Dilution	Analyst: CSB	NELAP
Analyses	Result	Qual	MRL	Units	Factor	Date Analyzed	Accredited
Lead	0.280	CF*	0.002	mg/L	20	2/20/2018	A

Lab ID: 1802626-03

Collection Date: 2/16/2018 6:24:00 AM

Matrix: DRINKING WATER

Client Sample ID: Girls RR 2nd Floor
Source: City Water
Sample Location: Girls RR 2nd Floor

Trace Metals by EPA 200.8 ICP-MS					Dilution	Analyst: CSB	NELAP
Analyses	Result	Qual	MRL	Units	Factor	Date Analyzed	Accredited
Lead	0.00570		0.0001	mg/L	1	2/20/2018	A

Qualifiers:

* Value exceeds Maximum Contaminant Level
E Value above quantitation range
J Analyte detected below quantitation limits
S Spike Recovery outside accepted recovery limits

B Analyte detected in the associated Method Blank
H Holding times for preparation or analysis exceeded
ND Not Detected at the Minimum Reporting Limit

Neilson Research Corporation

DATA FLAGS

B	Analyte detected in the associated method blank.
BA	BOD Alternative Calculation: The initial results performed by Standard Methods did not fall within parameters of the Standard Methods calculation. An alternate approved calculation was performed using the HACH method and the value reported is an estimated concentration.
C	Sample(s) does not meet NELAP/ORELAP sample acceptance criteria. See Case Narrative.
C1	Sample(s) does not meet NELAP/ORELAP sample acceptance criteria for temperature.
CF	Results confirmed by re-analysis.
CU	Cleanup performed as specified by method.
D1	The diesel elution pattern for the sample is not typical.
D2	The sample appears to be a heavier hydrocarbon range than diesel.
D3	The sample appears to be a lighter hydrocarbon range than diesel.
D4	Detected hydrocarbons do not have pattern and range consistent with typical petroleum products and may be due to biogenic interference.
D5	Detected hydrocarbons in the diesel range appear to be weathered diesel.
E	Estimated value.
ER	Elevated reporting limit due to matrix. Report limits (MDLs, MRLs & PQLs) are adjusted based on variations in sample preparation amounts, analytical dilutions, and percent solids, where applicable.
FC	Fecal Coliforms: Sample(s) received past 40 CFR Part 136 specified holding time. Results reported as estimated values.
G1	The gasoline elution pattern for the sample is not typical.
G2	The sample appears to be a heavier hydrocarbon range than gasoline.
G3	The sample appears to be a lighter hydrocarbon range than gasoline.
G4	Detected hydrocarbons in the gasoline range appear to be weathered gasoline.
HP	Sample re-analysis performed outside of method specified holding time.
HR	Sample received outside of method specified holding time.
HS	Sample analyzed for volatile organics contained headspace.
HT	At the client's request, the sample was analyzed outside of method specified holding time.
H	Analysis performed outside of method specified holding time.
J	Analyte detected below the Minimum Reporting Limit (MRL) and above the Method Detection Limit (MDL). The J flag result is an estimated value and the user should be aware that this data is of limited reliability.
L	Dissolved metals were not filtered within 15 minutes of collection per 40 CFR Part 136.
MI	Surrogate or Matrix Spike recovery is out of control limits due to matrix interference. Sample results may be biased.
N	See Case Narrative on page 2 of report.
Q	Closing continuing calibration verification (CCV) or laboratory control sample (LCS) exceeded high recovery limits, but associated samples are non-detect and the sample results are not affected. Data meets EPA/NELAP requirements.
R	Relative percent difference (RPD) is outside of the accepted recovery limits.
R1	Relative percent difference (RPD) is outside of the accepted recovery limits. However, analyses are not controlled on RPD values for sample concentrations that are less than the reporting limit.
R3	The relative percent difference (RPD) and/or percent recovery for the duplicate (DUP) or matrix spike (MS)/matrix spike duplicate (MSD) cannot be accurately calculated due to the concentration of analyte already present in the sample.
R4	Duplicate analysis failed due to result being at or near method reporting limit.
S	Surrogate and/or matrix spike recovery is outside of the accepted recovery limits. Sample results may be biased.
S1	Surrogate or matrix spike recovery is outside of control limits due to dilution necessary for analysis.
SC	Sub-contracted to another laboratory for analysis.
SP	Sample(s) were not collected per EPA Method 5035A protocols. The results are considered minimum values.
T	Toxicity Characteristic Leaching Procedure – Sample submitted contained < 0.5% solids. If the waste contains <0.5% dry solids, the liquid portion of the waste, after filtration, is defined as the TCLP extract.
#	Value exceeds regulatory level for TCLP contaminant.
X1	The motor oil elution pattern for the sample is not typical.
X2	The sample appears to be a heavier hydrocarbon range than motor oil.
X3	The sample appears to be a lighter hydrocarbon range than motor oil.
*	Value exceeds Maximum Contaminant Level or is outside the acceptable range.

CLIENT: Central Point Sch. Dist/Sams Valley Elem
Work Order: 1802626
Project: Lead Testing-D.O.

ANALYTICAL QC SUMMARY REPORT

TestCode: ICPMS_200.8_SCHOOL

Sample ID MB-40401	SampType: MBLK	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 2/20/2018	RunNo: 101570						
Client ID: ZZZZZ	Batch ID: 40401	TestNo: EPA 200.8	(EPA 200.8)	Analysis Date: 2/20/2018	SeqNo: 1530899						
Analyte	Result	MRL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Lead ND 0.000100

Sample ID LCS-40401	SampType: LCS	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 2/20/2018	RunNo: 101570						
Client ID: ZZZZZ	Batch ID: 40401	TestNo: EPA 200.8	(EPA 200.8)	Analysis Date: 2/20/2018	SeqNo: 1530900						
Analyte	Result	MRL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Lead 0.1018 0.000100 0.1 0 102 85 115

Sample ID 1802676-03AMS	SampType: MS	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 2/20/2018	RunNo: 101570						
Client ID: ZZZZZ	Batch ID: 40401	TestNo: EPA 200.8	(EPA 200.8)	Analysis Date: 2/21/2018	SeqNo: 1531439						
Analyte	Result	MRL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Lead 0.1037 0.000100 0.1 0.000836 103 70 130

Sample ID 1802676-03AMSD	SampType: MSD	TestCode: ICPMS_200.8	Units: mg/L	Prep Date: 2/20/2018	RunNo: 101570						
Client ID: ZZZZZ	Batch ID: 40401	TestNo: EPA 200.8	(EPA 200.8)	Analysis Date: 2/21/2018	SeqNo: 1531440						
Analyte	Result	MRL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual

Lead 0.09550 0.000100 0.1 0.000836 94.7 70 130 0.1037 8.20 20

Qualifiers: E Value above quantitation range H Holding times for preparation or analysis exceeded J Analyte detected below quantitation limits
 ND Not Detected at the Minimum Reporting Limit R RPD outside accepted recovery limits S Spike Recovery outside accepted recovery limits

